

INSTRUCTION SHEET



KEPCO An ISO 9001 Company.

**KIT
219-0642**

CHASSIS SLIDES KIT NO. 219-0642

1. DESCRIPTION

This Kepco KIT contains the components required to install chassis slides on a Kepco Model 0027633 power supply.

2. INSTALLATION OF CHASSIS SLIDES

2.1 MATERIAL REQUIRED (SEE TABLE 1.)

TABLE 1. MATERIAL REQUIRED

MATERIAL	LOCATION	QUANTITY
Slide, right, Kepco P/N 108-0530 (Jonathan P/N 15016458B-R)	Provided in this Kit	1
Slide, left, Kepco P/N 108-0529 (Jonathan P/N 15016458B-L)	Provided in this Kit	1
Screw 10-32 x 1/2 in. long, FLPO, stainless steel, Kepco P/N 101-0375	Provided in this Kit	14

2.2 PROCEDURE (SEE FIGURE 1)

CAUTION: Only use screws supplied in this KIT to mount chassis slides to Model 0027633 chassis. Use of any other screws, including either longer or shorter, may damage the unit.

1. Attach right slide section (P/N 108-0530) to right side of Model 0027633 chassis using seven screws (P/N 101-0375) (see Figure 1).
2. Attach left slide section (P/N 108-0529) to left side of Model 0027633 chassis using seven screws (P/N 101-0375)

CAUTION:
USE ONLY SCREWS SUPPLIED IN THIS KIT TO
INSTALL SLIDES. FAILURE TO USE HARDWARE
SPECIFIED OR USING SHORTER OR LONGER
SCREWS CAN DAMAGE THE UNIT.




FIGURE 1. INSTALLING CHASSIS SLIDES ON MODEL 0027633 POWER SUPPLY

KEPCO, INC. ● 131-38 SANFORD AVENUE ● FLUSHING, NY. 11355 U.S.A. ● TEL (718) 461-7000 ● FAX (718) 767-1102
www.kepcopower.com ● email: hq@kepcopower.com